Ascentcare™
Dental Products

Instruction and care guides for:

Bite Buddy™
Mouth Prop
BBL, BBM, BBS

Tongue Tamer™
Tongue Guard
TTLGR, TTLGL, TTSMR, TTSML

Saliva Sidekick™
Saliva Ejector Tube Holder
SSKR, SSKL

Placement Pal™
Accessory Placement Instrument
BBPP

Hero Holder™
Bite Buddy System Organizational Tray
BBHH

Bite Buddy™
Mouth Prop

1. When sterilized and ready to use, floss may be tied around the Bite Buddy for quick retrieval in the event of dislodgement or aspiration.
2. Place in patient's mouth over back molars using Placement Pal or index finger.
3. Do not place over front teeth.
4. To remove, see Placement Pal instructions or remove by hooking with index finger and gently pulling outwards.

Tongue Tamer™
Tongue Guard

1. When sterilized and ready to use, tie dental floss to the hole in the tongue guard to allow for quick retrieval in the event of dislodgement or aspiration.
2. Slide the Tongue Tamer into the magnetic track on the Bite Buddy, ensuring that it is secure in the track.
3. Position the Bite Buddy with the Tongue Tamer into the mouth according to the Bite Buddy instructions.
4. The Tongue Tamer can be removed with the Bite Buddy or slid from the track while the Bite Buddy remains in the patient's mouth. Simply hook your index finger around the tongue guard and pull it free from the magnetic track.

Saliva Sidekick™
Saliva Ejector Tube Holder

1. When sterilized and ready to use, position your saliva ejector tube under the tube holder and press it up into the holder until it is secure.
2. Slide the Saliva Sidekick into the magnetic track on the Bite Buddy, ensuring that it is securely in place.
3. Position the Bite Buddy with the Saliva Sidekick into the mouth according to the Bite Buddy instructions.
4. The Saliva Sidekick can be removed with the Bite Buddy or slid from the track while the Bite Buddy remains in the patient's mouth. Without pulling upwards or downwards, gently slide the saliva ejector tube with the Saliva Sidekick straight out of the Bite Buddy track.

Love your Saliva Sidekick?



Check out our website to see the latest Bite Buddy accessory: the Lighted Saliva Sidekick™

Placement Pal™
Accessory Placement Instrument

To insert the Bite Buddy:

1. When sterilized and ready to use, slide the Placement Pal into the Bite Buddy track, ensuring it is secure.
2. Hold the Placement Pal with your index finger and thumb to place the Bite Buddy in the proper position according to the Bite Buddy instructions.
3. Gently slide the Placement Pal out of the track to remove from the Bite Buddy.

To remove the Bite Buddy:

1. The Placement Pal cannot be used while other accessory instruments are attached to the Bite Buddy. Remove the other instruments according to their instructions before using the Placement Pal.
2. While the Bite Buddy is in the patient's mouth, slide the Placement Pal into the magnetic track on the Bite Buddy, ensuring it is secure.
3. Gently pull the Placement Pal and the Bite Buddy out of the mouth together.



Sterilization Instructions:

Remove from packaging and sterilize all products according to these instructions before use.

Sterilization should be performed using chemical methods or via autoclave. Insufficient water to the autoclave could cause parts to melt. Cold Sterilization is the recommended method of sterilization.

Cold Sterilization:

1. As soon as possible after use, soak and rinse your Ascentcare instruments in warm water to remove any blood or tissues.

2. Select a liquid chemical agent registered with the EPA as a "sterilant."

3. Follow the sterilization instructions of your sterilant liquid. Most cold sterilization solutions render instruments sterile only after 10-hour immersion.

4. Rinse, dry and store in a sterile container if not used immediately.



Hero Holder™
Organizational Tray

The Hero Holder neatly organizes your Bite Buddies and accessories. Slide the accessories into the slots as pictured above. The accessories will be magnetically held into place. Sterilize via autoclave or chemical methods according to sterilzation instructions in this pamphlet.

Ascentcare's products are protected by US Patents.
For patent information, visit:
www.ascentcaredental.com/patents

The Bite Buddy and all accessories are latex free.

As with all products containing magnets, please take care to avoid inadvertently deactivating such devices as pacemakers, ICD's and other implanted medical devices.

Be sure to take all necessary precautions to avoid dislodgement or aspiration by following the instructions in this pamphlet.

All Ascentcare's Products are proudly made in the USA, in our hometown of Spring Lake, MI.

Ascentcare™
Dental Products
PO BOX 185 Spring Lake, MI 49456
(800) 500-0533
www.ascentcaredental.com